


Product Specification Sheet

by ChraWork a Gimium™ Ecosystem

2110120343

LC COLUMN QUALITY CONTROL PERFORMANCE		
COLUMN INFORMATION SERIAL NUMBER : 2110120343 PART NUMBER : E07219-C4126 COLUMN TYPE : ChraSil Inert HILIC, 150 x 2.1 mm, 3.5um LOT : B08004		TEST CONDITIONS MOBILE PHASE : 0.1% TFA COLUMN FLOW RATE : 1.0 mL/min COLUMN FLOW RATE : 1.0 mL/min LINEAR VELOCITY : 0.1 cm/s TEMPERATURE : 30°C INJECTION VOLUME : 10 µL
TEST VALUES THEORETICAL PLATES : 23281 SELECTIVITY : 1.63 USP TAILING FACTOR : 1.06 (@ 5% Peak Height)	SPECIFICATIONS MIN : 18200 RANGE : 1.60 - 1.72 RANGE : 0.98 - 1.25 k' : 2.09	Peak 1. 2. 3. 4.
 Image not found or type unknown		
MATERIAL SAFETY DATA SHEETS ARE AVAILABLE		
ChraReport at https://chracert.com/2110120343		

Revision #2

Created 22 October 2021 01:57:38 by Admin

Updated 22 October 2021 01:59:03 by Admin