

# C0911-A8029

|                     |                                 |
|---------------------|---------------------------------|
| Description         | ChraSil A100 C8, 75 x 4.6 mm, 5 |
| Product Type        | HPLC Column (Liquid Chromatog   |
| For Use with        | HPLC Instrument                 |
| Manufacturer        | Chratec                         |
| Internal Diameter   | 4.6                             |
| Column Length       | 75                              |
| Particle Size       | 5.0                             |
| Modification        | C8                              |
| Pressure Stability  | 400 bar                         |
| Base Material       | Silica, spherical, porous       |
| Packing brand       | ChraSil                         |
| USP L Code          | L7                              |
| pH Range#           | 2 - 8                           |
| Carbon Load#        | 12.5 %                          |
| Endcap              | + (Yes)                         |
| Pore Size           | 100Å                            |
| Surface Area(m2/g)  | 350                             |
| Operating Temp (°C) | 20 - 70                         |
| Column Group#       | A100                            |
| Part Number         | C0911-A8029                     |

|              |   |
|--------------|---|
| Installation | Please see instrument manual for  |
| Replacement  | Beyond routine maintenance  |
| Care and Use | <a href="https://chratesc.com/mag/CaU.pdf">https://chratesc.com/mag/CaU.pdf</a> |
| Warranty#    | See ChraContact at <a href="https://chratesc.com">https://chratesc.com</a>      |
| ChraSil URL# | C0911-A8029 at <a href="https://chrasil.com">https://chrasil.com</a>            |

Revision #1

Created 11 November 2021 02:31:40 by Admin

Updated 11 November 2021 02:31:54 by Admin