

C0926-A1025

| | |
|--------------------|--|
| Description | ChraSil HPLC Column |
| Manufacturer | Chratec |
| Internal Diameter | 4.6 |
| Column Length | 30 |
| Particle Size | 10.0 |
| Modification | C4 |
| Pressure Stability | 400 bar |
| Base Material | Silica, spherical, porous |
| Packing brand | ChraSil |
| USP L Code | L26 |
| Part Number | C0926-A1025 |
| pH Range# | 2 - 8 |
| Carbon Load# | - (See https://chratesil.com) |
| Endcap | # |
| Pore Size | - |
| Surface Area(m2/g) | - |
| Column Group# | - |
| ChraSil URL# | C0926-A1025 at https://chratesil.com |
| Installation | Please see instrument manual |
| Replacement | Beyond routine maintenance |

| | |
|--------------|---|
| Care and Use | https://chrates.com |
| Warranty# | See ChraContact at |

Revision #1

Created 29 July 2021 04:26:13 by Admin

Updated 29 July 2021 04:26:20 by Admin